



Evaluation of Cefoxitin Disc Diffusion and Chromogenic Agar in the Detection of Methicillin Resistant *Staphylococcus aureus*

A. I. Umar ^{a*}, I. Garba ^a and A. M. Ganau ^a

^a Department of Medical Microbiology, Faculty of Medical Laboratory Science, Usmanu Danfodiyo University Sokoto, Nigeria.

Authors' contributions

This work was carried out in collaboration among all authors. All the authors contributed in the study design data analysis and proofreading the manuscript. All authors read and approved the final manuscript.

Article Information

DOI: 10.9734/SAJRM/2023/v15i1278

Open Peer Review History:

This journal follows the Advanced Open Peer Review policy. Identity of the Reviewers, Editor(s) and additional Reviewers, peer review comments, different versions of the manuscript, comments of the editors, etc are available here: <https://www.sdiarticle5.com/review-history/96951>

Original Research Article

Received: 19/10/2022
Accepted: 29/12/2022
Published: 20/01/2023

ABSTRACT

Methicillin resistant *Staphylococcus aureus* (MRSA) has a worldwide distribution and is an important cause of clinical and epidemiological problems. The aim of this study was to evaluate the usefulness of some phenotypic methods for the detection of methicillin resistant *S. aureus* in clinical laboratories. The study is cross sectional. A total of 93 *S. aureus* isolates were tested using cefoxitin disk diffusion (CDD) and oxacillin resistance screening agar base (ORSAB) with reference to *mecA* gene PCR. Of the 93 isolates, CDD test showed 34 were MRSA, while ORSAB recorded 42. *MecA* gene was detected by PCR in 34 of the isolates. The CDD showed 97.1% sensitivity and 98.3% specificity and therefore superior to ORSAB with sensitivity 97.1% and specificity 84.7%. The cefoxitin disk test required no special test conditions and can improve the reliability of routine tests for the detection of MRSA. CDD test can thus be used as a cheap and reliable alternative to PCR for the detection of MRSA in resources limited settings.

*Corresponding author: Email: uiasiya@yahoo.com;

Keywords: Cefoxitin disk; Chromogenic agar; *mecA*; MRSA; *Staphylococcus aureus*.

1. INTRODUCTION

Staphylococcus aureus is one of the important pathogens in many countries causing infection in hospitals and the community. It causes a variety of diseases, ranging in severity from boils and furuncles to more serious diseases such as septicaemia, pneumonia and endocarditis [1]. "Virulent strains of these bacteria are responsible for the majority of Hospital Acquired Infections (HAIs) and can cause severe disease that can be fatal" [2]. "Since the emergence of *S. aureus* strains with resistance to methicillin in 1961 [3], it has become a well-known aetiological agent for a wide variety of infections. Methicillin-resistant *S. aureus* (MRSA) infections have become a common problem in hospital and community-acquired infections" [4,5,6]. "Rapid and accurate identification of MRSA is important in guiding clinicians choose appropriate antibiotics to treat and prevent spread of these strains. Many methods exist for detection of MRSA, and they include disk diffusion method which is the method used by most laboratories, use of chromogenic agars, E-test, the Minimal Inhibitory Concentration (MIC) by broth dilution or E-test, use of automated machines for identification and more recently molecular methods for detection of *mecA* gene" [7]. "The *mecA* gene is highly conserved among the *Staphylococci* species, and so the importance of PCR in assay of the *mecA* gene as "gold standard" for the detection of methicillin resistance in *Staphylococci* is well established" [8,9,10]. "However, the use of PCR is not practical for routine diagnostic use in clinical laboratories due to cost and need for technical and experienced staff" [7]. In view of these, there is need for a reliable and simple test to detect MRSA easily in routine clinical laboratories.

2. MATERIALS AND METHODS

Bacterial isolates used in this study were obtained from Usmanu Danfodiyo University Teaching Hospital, Specialist Hospital and Maryam Abacha Women and Children Hospital all in Sokoto, North Western Nigeria. All isolates were identified by standard microbiological methods including colony morphology, Gram staining, catalase test, coagulase test using both slide and tube methods and deoxyribonuclease test [11,12]. All isolates were examined for methicillin resistance by cefoxitin disk (30µg) diffusion test (CDD), Oxacillin resistance screen

agar base (ORSAB) and Polymerase Chain Reaction (PCR).

Inoculum for Antibiotic Susceptibility Testing was prepared by the direct colony suspension method recommended by the Clinical Laboratory Standards Institute for testing *Staphylococci* for potential methicillin resistance [13].

2.1 Cefoxitin Disk Diffusion Test

Cefoxitin disk 30µg (Oxoid, UK) was used on Mueller Hinton agar (Oxoid, UK). The inoculum was standardized by comparing with 0.5 McFarland turbidity standard and for the 0.5 McFarland standard, the absorbance at a wavelength of 625 nm was 0.09, the agar plates inoculated, inverted and incubated at 35°C for 24h. After the incubation, the zone of inhibition was measured using a metre rule (against transmitted light) and the results (zones ≤21mm indicate resistance) interpreted using the CLSI [14] guidelines.

2.2 Oxacillin Resistance Screening Agar Base (ORSAB) Test

Chromogenic media was prepared according to manufacturer's instruction. A 0.05ml aliquot of broth was plated onto ORSA plates and incubated at 35°C in ambient air. Growth on plates was evaluated after 18 hours of incubation as recommended by the manufacturer. Deep blue colonies on ORSAB were classified as MRSA. Coagulase Negative *Staphylococci* produce white colonies on ORSA (Oxoid, England). Any bacterial growths after 24 h, resulting in intense blue colonies were indicative of resistance to methicillin as described by Simor et al. [15]. Results obtained were compared to those obtained by cefoxitin disk diffusion test.

2.3 Molecular Analysis of the *S. aureus* Isolates

DNA extraction for amplification was done using the method of Stephenson [16] but with slight modification. Genomic DNA was prepared for all polymerase chain reaction (PCR) methods. Two to three colonies of 18 hours cultured *S. aureus* was added to 100 µL of a nuclease and ligase inhibitor-free water for extraction of PCR-ready template DNA in an eppendorf tube. This was vortexed for 15 seconds and incubated at 95 °C for 30 min and then centrifuged at 13,000 rpm for 5 min to allow purification of the DNA as the

heavier molecules were settled at the bottom and the DNA was suspended in the supernatant. The supernatant (70 µL) was transferred into a new tube and stored at -20 °C as described by Stephenson [16].

2.4 Duplex PCR Reaction Mix of 16S rRNA and *mecA* Gene from the Extracted DNA

RNA (16S rRNA) region with the following primer sequence was used as the quality control for the detection of extracted DNA. 27F (5'-AGAGTTTGATCMTGGCTCAG-3')907R (5'-CCGTC AATTCMTTTRAGTTT-3') as described by Jiang et al., [17]. The 25µL volume of PCR reaction mixture used, contained 3 µL of genomic DNA, 12.5 µL of Mastermix (BioLabs New England) containing; 25 units/ml Taq DNA polymerase, 32 mM (NH₄)₂SO₄, 10mM Tris HCl, 0.05% Tween 20, 1.5 mM MgCl₂, 50 Mm KCL 5% Glycerol and dNTPs: dATP, dCTP, dGTP, dTTP, then 5.5 µL nuclease free H₂O, 1 µL each of 16S rRNA primers (27F and 907R) and 1µL each of *mecA* primers (forward and reverse). *mecA* gene was amplified with the following primers: *mecA*-f: (5'-AAAATCGATGGTAAAGGTTGGC-3'); *mecA*-r: (5'-AGTTCTGCAGTACCGGATTTGC-3') with 533bp. Negative control contains all material except template DNA, distilled water was added to the negative control. DNA amplification was carried out for 40 cycles according to the following protocol: denaturation at 94 °C for 30 s, annealing at 55 °C for 30 s, and extension at 72 °C for 1 min with a final extension at 72 °C for 5 min. The PCR products were then analyzed in 2% (w/v) agarose gels stained with ethidium bromide and visualized under UV light as described by Murakami et al. [18].

2.5 Gel Electrophoresis

The electrophoresis chamber was filled with the running buffer (1x TBE) and casting tray was then placed in the chamber. The gel was immersed in the buffer. 2µL of loading dye was

added to each of the PCR products. The amplicon (10µL) was loaded into each well. A 1kb plus molecular weight marker- *Hind* III Lambda DNA ladder was then loaded into the first and the last well as standard for estimating the size of the resulting DNA fragment. The electrophoresis chamber was then connected to the power source and the DNA run at 160V for 45 min as described by Murakami et al. [18]. The DNA bands (976 bp for 16s and 533bp for *mecA*) were then viewed by illumination with UV light and images recorded by photography (Biorad imager).

3. RESULTS

Table 1 shows comparison between two methods; Disc diffusion and ORSAB chromogenic method for the detection of Methicillin resistance among the isolates. Of the 93 isolates, 34 with a zone of inhibition ≤21mm (36.6%) showed Methicillin resistance using the disc diffusion method and 42(45.2%) showed Methicillin resistance using the ORSAB chromogenic medium. All except one of the isolates that were cefoxitin resistant demonstrated resistance using the ORSAB Medium. Table 2 shows the comparison between PCR and Cefoxitin disc diffusion. Using the Chi square test, the variation between the two tests is statistically significant with a P value of 0.000.

Table 3 shows the comparison between the PCR and ORSAB methods and the variation between the two tests is also statistically significant with a P value of 0.000.

The statistical inferences in which PCR was used to compare the methods using the two by two contingency table are shown in Table 4, in which the Cefoxitin DD method was found to be more sensitive and specific with 97.1% and 98.3% than the ORSAB Method with 97.1% and 84.7% respectively. The Cefoxitin DD test was also found to have a higher positive predictive value (97.1%).

Table 1. Comparison between Cefoxitin, Disc Diffusion and Oxacillin Resistance Screening Agar for the detection of Methicillin Resistant *Staphylococcus aureus* (MRSA) in the study centers

Method		ORSAB		Total	P value
		Sensitive	Resistant		
Cefoxitin DD Test	Sensitive	33	1	34	0.000
	Resistant	9	50	59	
Total		42	51	93	

$$\chi^2 = 58.284$$

Table 2. Comparison of mecA gene detection by PCR and Disk Diffusion using cefoxitin in the detection of MRSA isolates

Method		Polymerase Chain Reaction		Total	P value
		Sensitive	Resistant		
Cefoxitin DD Test	Sensitive	58	1	59	0.000
	Resistant	1	33	34	
Total		59	34	93	

$\chi^2 = 84.577$

Table 3. Comparison of mecA gene detection by PCR and ORSAB in the detection of MRSA isolates

Method		Polymerase Chain Reaction		Total	P value
		Sensitive	Resistant		
ORSAB	Sensitive	50	1	51	0.000
	Resistant	9	33	42	
Total		59	34	93	

$\chi^2 = 58.284$

Table 4. Evaluation of Cefoxitin DD and ORSAB using PCR as gold standard for the detection of Methicillin resistance in the *Staphylococcus aureus* isolates

Statistics	Type of method evaluated	
	Cefoxitin DD Test	ORSAB
Sensitivity	97.1	97.1
Specificity	98.3	84.7
Positive Predictive Value	97.1	78.6
Negative Predictive Value	98.3	98.0




Fig. 1. Results of 16s (976 bp product) and mecA gene Polymerase Chain Reaction (533 bp product)

Key: L: 1kb plus molecular ladder; -ve: negative control (PCR premix only); +ve: positive control (*S. aureus* ATCC 25923); 3: 35s; 4: 2s; 5: 15u; 6: 17s

4. DISCUSSION

This study showed that 34 (36.6%) of the isolates obtained were methicillin resistant by the cefoxitin disc diffusion method. While oxacillin resistance screening agar base, a chromogenic agar, on the other hand, detected 42 (45.2%)

MRSA. Eight more MRSA than those detected by cefoxitin DD test were captured by ORSAB and is corroborated by other findings [19,20,21]. This is probably because the media detect key microbial enzymes as diagnostic markers for pathogens through the use of chromogenic substrates incorporated into the agar base

allowing direct colony color base identification of MRSA, as against the disc diffusion which relies on antibiotic inhibition of the bacteria. This finding is however in contrast with findings of Anand et al. [22] in India, who reported that the cefoxitin DD captured more isolates with methicillin resistance than the ORSAB method. This study compared the performance of ORSAB and the Cefoxitin 30 µg disk diffusion, and it was observed that there is a statistical significant difference in the two methods ($P < 0.05$). In the past decades the performance of proprietary Chromogenic media for the early identification of MRSA has improved dramatically. The ORSAB can be used by direct inoculation or from a primary culture grown on other types of media. MRSA colonies appear bluish; this makes them very easy to see against the light coloured opaque background of the media. Results can be obtained within 18 hours, this gives a rapid turnaround time when compared with standard culture that will require on the minimum of 48 hours. This enables for rapid patient care and reduces the risk of nosocomial spread of infections [23]. However, to determine the most effective method of detecting methicillin resistance between the two methods, molecular detection of the genes responsible for methicillin resistance in *S. aureus* was done.

Polymerase chain reaction has been considered as the gold standard for determining resistance genes as these genes are highly conserved among Staphylococcal species and was therefore used in this study. Of the 93 isolates analysed to detect the presence or otherwise of *mecA* gene, 34(36.6%) were found to harbour the gene. This in agreement with findings of Angela et al. [24]; Adetayo et al. [25]; Murakami et al. [18] and Hawraa and May [26] who all detected *mecA* gene in *S. aureus* isolates. "PCR based assays are considered as the gold standard for the detection of MRSA due to the heterogeneous resistance by various phenotypic detection methods displayed by many clinical isolates. Genotypic methods are more accurate in detecting MRSA as compared to conventional susceptibility methods" [26].

The sensitivity and specificity for detection of methicillin resistance using PCR as the gold standard in this study was 97.1% and 98.3% for cefoxitin DD and 97.1% and 84.7% for ORSAB. This suggests that cefoxitin DD test both sensitive and more specific in accurate detection of methicillin resistance than ORSAB. This is in agreement with findings of Rakesh et al. [20] who

reported sensitivity and specificity of 96.7% and 100% for cefoxitin as against 95.1% and 100% for ORSAB in India, Uzun et al. [13] with 98.3% and 100% for cefoxitin DD against 96.7% and 81.6% for ORSAB in Turkey and Boutiba et al. [21] with 100% and 96.5% for cefoxitin DD against 99.1% and 90.4% for ORSAB in Tunisia. All of these findings indicate that cefoxitin DD method can be used as a reliable conventional, simple and cheap alternative to PCR for detection of MRSA especially in resource constraint settings.

5. CONCLUSION

The cefoxitin disk test required no special test conditions and can improve the reliability of routine tests for the detection of MRSA. CDD test can thus be used as a cheap and reliable alternative to PCR for the detection of MRSA in resources limited settings.

CONSENT

Informed consent was sought from all the enrollees before collecting their samples.

ETHICAL APPROVAL

Ethical approval was obtained from the institutional ethical review committees of the hospitals before the commencement of the study.

ACKNOWLEDGEMENTS

We wish to acknowledge the contribution of the Laboratory staff of the hospitals where the samples for the study was sought, also worthy of mention are the study population and staff of laboratory, Department of Medical Microbiology, faculty of Medical Laboratory Science, Usmanu Danfodiyo University Sokoto, North-western Nigeria

COMPETING INTERESTS

Authors have declared that no competing interests exist.

REFERENCES

1. Lowy FD. Staphylococcus aureus infections. N Engl J Med. 1998;339(8): 520-32.
DOI: 10.1056/NEJM199808203390806, PMID 9709046.

2. Queck SY, Khan BA, Wang R, Bach TH, Kretschmer D, Chen L et al. Mobile genetic element-encoded cytolysin connects virulence to methicillin resistance in MRSA. *PLOS Pathog.* 2009;5(7): e1000533. DOI: 10.1371/journal.ppat.1000533, PMID 19649313.
3. Jevons MP. Celbenin-resistant staphylococci. *Br Med J.* 1961;1:124-5.
4. Kopp BJ, Nix DE, Armstrong EP. Clinical and economic analysis of methicillin-susceptible and -resistant *Staphylococcus aureus* infections. *Ann Pharmacother.* 2004;38(9):1377-82. DOI: 10.1345/aph.1E028, PMID 15266044.
5. Lodise TP, McKinnon PS. Clinical and economic impact of methicillin resistance in patients with *Staphylococcus aureus* bacteremia. *Diagn Microbiol Infect Dis.* 2005;52(2):113-22. DOI: 10.1016/j.diagmicrobio.2005.02.007, PMID 15964499.
6. Nixon M, Jackson B, Varghese P, Jenkins D, Taylor G. Methicillin-resistant *Staphylococcus aureus* on orthopaedic wards: Incidence, spread, mortality, cost and control. *J Bone Joint Surg Br.* 2006;88(6):812-7. DOI: 10.1302/0301-620X.88B6.17544, PMID 16720779.
7. Kaur N, Prasad R, Varma A. Detection of methicillin resistant *Staphylococcus aureus* (MRSA) by automated and manual methods. *Int J Pharm Biol Sci.* 2013;4(2):(B) 534 – 540.
8. Ekrami A, Samarbafzadeh AR, Alavi M, Kalantar E, Hamzelo F. Prevalence of methicillin resistant *Staphylococcus* species isolated from burn patients in a burn center, Ahvaz, Iran. *Jundishapur J Microbiol.* 2010;3(2):84-91.
9. Mathews AA, Thomas M, Appalaraju B, Jayalakshmi J. Evaluation and comparison of tests to detect methicillin resistant *S. aureus*. *Indian J Pathol Microbiol.* 2010; 53(1):79-82. DOI: 10.4103/0377-4929.59189, PMID 20090228.
10. Shariati L, Validi M, Tabatabaiefar MA, Karimi A, Nafisi MR. Comparison of real-time PCR with disk diffusion, agar screen and E-test methods for detection of methicillin-resistant *Staphylococcus aureus*. *Curr Microbiol.* 2010;61(6): 520-4. DOI: 10.1007/s00284-010-9647-9, PMID 20405128.
11. Garcia LS, Isenberg HD. *Clinical Microbiology Procedures handbook.* 3rd ed. American society of microbiology press. Washington, DC; 2010.
12. United Kingdom Standards for Microbiology Investigations Bacteriology test procedures. Issued by the Standards Unit, Pub. health Engl. 2014;(Issue no 3):8-11.
13. Clinical and Laboratory Standards Institute (CLSI) performance standards for antimicrobial disk susceptibility tests; approved standard. 11th ed. 2012;11-3.
14. Clinical and Laboratory Standards Institute (CLSI) performance standards for antimicrobial 'susceptibility testing; twenty-fourth informational supplement (M100-S24)',. Valley Road: West, Suite 2500 Wayne, PA 19087. USA. 2014;950:68-75.
15. Simor AE, Goodfellow J, Louie L, Louie M. Evaluation of a new medium, oxacillin resistance screening agar base, for the detection of methicillin-resistant *Staphylococcus aureus* from clinical specimens. *J Clin Microbiol.* 2001;39(9): 3422. DOI: 10.1128/JCM.39.9.3422.2001, PMID 11558472.
16. Stephenson FH. *Calculations for molecular biology and biotechnology: A guide to mathematics in the laboratory.* Elsevier Science; 2003.
17. Jiang H, Dong H, Zhang G, Yu B, Chapman LR, Fields MW. Microbial diversity in water and sediment of lake Chaka, an athalassohaline lake in north western China. *Appl Environ Microbiol.* 2006;72(6):3832-45. DOI: 10.1128/AEM.02869-05
18. Murakami K, Minamide W, Wada K, Nakamura E, Teraoka H, Watanabe S. Identification of methicillin-resistant strains of *Staphylococci* by polymerase chain reaction. *J Clin Microbiol.* 1991;29(10): 2240-4. DOI: 10.1128/jcm.29.10.2240-2244.1991, PMID 1939577.
19. Uzun B, Karataş Şener AG, Güngör S, Afşar I, Yüksel Ergin O, Demirci M. Comparison of cefoxitin disk diffusion test, automated system and chromogenic medium for detection of methicillin resistance in *Staphylococcus aureus* isolates. *Mikrobiyol Bul.* 2013;47(1):11-8. DOI: 10.5578/mb.4304, PMID 23390898.

20. Rakesh KP, Ashoka M, Bandana M, Nirupama C. Evaluation of Genotypic and phenotypic Methods for Detection of methicillin resistant Staphylococcus aureus in a tertiary care Hospital of Eastern Odisha. J Clin Diagn Res. 2016;10(2): 19-21.
21. Boutiba-Ben Boubaker I, Ben Abbes R, Ben Abdallah H, Mamlouk K, Mahjoubi F, Kammoun A et al. Evaluation of a cefoxitin disk diffusion test for the routine detection of methicillin resistant Staphylococcus aureus. Clin Microbiol Infect. 2004;10(8):762-5.
DOI: 10.1111/j.1469-0691.2004.00919.x, PMID 15301683.
22. Anand KB, Agrawal P, Kumar S, Kapila K. Comparison of cefoxitin disc diffusion test, oxacillin screen agar, and PCR for mecA gene for detection of MRSA. Ind J Med Microbiol. 2009;27(1):27-9.
DOI: 10.1016/S0255-0857(21)01748-5, PMID 19172055.
23. Veenemans J, Verhulst C, Punselie R, van Keulen PHJ, Kluytmans JAJW. Evaluation of brilliance MRSA 2 agar for detection of methicillin-resistant Staphylococcus aureus in clinical samples. J Clin Microbiol. 2013;51(3):1026-7.
DOI: 10.1128/JCM.02995-12, PMID 23284023.
24. Angela B, Aparecida C, Miranda C, Flores C, de Filippis I, Maysa M. Clementino methicillin- and vancomycin-resistant Staphylococcus aureus in health care workers and medical devices. Braz J Path Lab Med. 2015;51(3):143-52.
25. Adetayo TO, Deji-Agboola AM, Popoola MY, Atoyese TJ, Egberongbe KY. Prevalence of MRSA from clinical specimens in Ibadan Nigeria. Int J Eng Sci. 2014;3(9):1-11.
26. Hawraa AM, May TF. Detection of vanA gene of vancomycin resistant Staphylococcus aureus by PCR technique. Int J Adv Res. 2014;2(7):209-16.

© 2023 Umar et al.; This is an Open Access article distributed under the terms of the Creative Commons Attribution License (<http://creativecommons.org/licenses/by/4.0>), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Peer-review history:

The peer review history for this paper can be accessed here:
<https://www.sdiarticle5.com/review-history/96951>